



Laboratory
Accreditation
Bureau

Certificate of Accreditation

ISO/IEC 17025:2005

Certificate Number L2060-1

Scale Services & Sales
28 Crown Steel Drive, Unit 7
Markham, ON L3R 9Y1

has met the requirements set forth in L-A-B's policies and procedures, and all requirements of ISO/IEC 17025:2005
"General Requirements for the competence of Testing and Calibration Laboratories." This accreditation
demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(refer to joint ISO-ILAC-IAF Communiqué dated January 2009).

Accreditation valid through September 8, 2012

R. Douglas Leonard, Jr., Managing Director
Laboratory Accreditation Bureau

Scope of Accreditation For Scale Services & Sales

28 Crown Steel Drive, Unit 7
Markham, ON L3R 9Y1
M. (Roger) Sinnarajah
905-940-8320

In recognition of a successful assessment to ISO/IEC 17025:2005, accreditation is granted to **Scale Services & Sales** to perform the following Calibrations:

Accreditation granted through: **September 8, 2012**

Calibration

Mass – Scales and Balances

Calibration Parameter/Equipment	Range	Calibration and Measurement Capability(+/-)	Remarks
Laboratory Balances and High Precision Scales Display Resolution: 0.0001 g	(0 to 210) g	0.00018 g	ASTM E617 Class 1 Weights, Canadian Weights & Measures Act & Regulations, NIST Handbook 44 and Scale Services & Sales Procedures
0.001 g	(0 to 410) g	0.0015 g	
0.01 g	(0 to 4100) g	0.013 g	
Bench Scales and Counting Scales Display Resolution: 0.005 g	(0 to 500) g	0.0053 g	ASTM E617 Class 1 Weights, ASTM E617 Class 4 Weights, Canadian Weights & Measures Act & Regulations, NIST Handbook 44 and Scale Services & Sales Procedures
0.01 g	(0 to 5000) g	0.016 g	
0.01 g	(0 to 2000) g	0.013 g	
0.02 g	(0 to 2000) g	0.021 g	
0.05 g	(0 to 5000) g	0.053 g	
0.1 g	0 g to 12 kg	0.084 g	
0.1 g	(12 to 32) kg	0.091 g	
0.2 g	0 g to 12 kg	0.16 g	
0.5 g	(0 to 25) kg	0.57 g	
1 g	(0 to 50) kg	1.2 g	
5 g	(0 to 50) kg	4.1 g	
10 g	(0 to 100) kg	8.2 g	

Calibration Parameter/Equipment	Range	Calibration and Measurement Capability(+/-)	Remarks
Floor Scales Display Resolution: 20 g	(0 to 200) kg	17 g	OIML R 111 Class M1 Weights, Canadian Weights & Measures Act & Regulations, NIST Handbook 44 and Scale Services & Sales Procedures
50 g	(0 to 200) kg	41 g	
100 g	(0 to 300) kg	82 g	
200 g	(0 to 500) kg	164 g	
0.5 kg	(0 to 2500) kg	0.6 kg	
1 kg	(0 to 5000) kg	1.2 kg	

Calibration and Measurement Capability (CMC) represents expanded uncertainties at approximately a 95% confidence level using a coverage factor of k=2.

Notes:

- 1) Laboratory offers calibration services at the laboratory's own facilities and at the client or other agreed upon facilities.

Approved by: _____



R. Douglas Leonard
Chief Technical Officer

Date: October 17, 2011